

E53.A Series Piezo Controller User Manual

Version: V1.0



This document describes the following products:

■ E53.A Open loop 1 channel

DECLARATION

Declaration!

This user manual is a integrated user manual of the E53.A series piezoelectric controller. Please read this user manual carefully before using this controller. Follow the instructions in the manual during use. If there is any problem, please contact us for technical support. If you do not follow this manual or disassemble and modify the product yourself, the company will not be liable for any consequences arising therefrom.

Please read the following to avoid personal injury and to prevent damage to this product or any other product connected to it. In order to avoid possible hazards, this product can only be used within the specified range.

Notice!

Do not touch any exposed ends of the product and its accessories.

There is high voltage inside. Do not open the case without permission.

Do not connect or disconnect input, output, or sensor cables with power on.

Please keep surface of E53.A clean and dry, don't operate in humid or static environment.

After use, output voltage should be cleared to zero before turning off the controller switch, such as switching the servo state to the open-loop state.

Danger!

The piezoelectric power amplifier described in this manual is a high-voltage device capable of outputting high currents, which can cause serious or even fatal damage if not used properly.

It is strongly recommended that you do not touch any parts that connect to the high voltage output. Special Note: If you connect it with other products in addition to our company, please follow the general accident prevention procedures.

Operating the high-voltage amplification requires training professional operators.

Warning!

If the voltage exceeds the PZT's tolerable range, it will cause permanent damage to the PZT. Before adding voltage to the PZT poles, it must be ensured that the positive and negative poles of the PZT are connected correctly and the operating voltage is within the allowable range of this PZT.

Cautious!

E53.A housing should be installed on a horizontal surface in an area with a 3CM air flow area to prevent internal convection in the vertical direction.

Insufficient airflow can cause equipment to overheat or premature instrument damage.

Contents

1. Introduction	. 2
1.1 Features	. 2
1.2 Applicaitons	. 2
1.3 Order information	. 2
2.Driving Principle	. 2
3.Appearance	
4.Power Calculation	. 3
5.Panel Introduction	. 4
6.Drawings	. 4
7.Notes and suggestions	. 5
8.Contact us	. 5



1.1 Features

- ▶ 1 channel small size
- ▶ 24V (20~30V) 1.5A 36W
- ▶ Peak current 1A
- ▶ Ave current 60mA
- ▶ Unload bandwidth 10KHz
- Output short circuit protection

1.2 Applicaitons

- Driving piezo actuators
- Driving piezo objective scanners


1.3 Order information

▶ E53.A——Open loop

Accept customized according to requirements:

- ① 12bit gain/-20 ~ 120V output voltage (default)
- 2 15bit gain/-20 ~ 150V output voltage











4.Power Calculation

• Average output (Sine wave operation mode)

Pa ≈ Upp • Us • f• Cpiezo

Pa=Average output[W]


Cpiezo=Piezo actuator capacitance[F]

Upp=Peak and peak drive voltage [V]

f=Operating frequency of the sine wave[Hz]

 $Us=Drive\ voltage[V]\ (\ (Vs+)\ -\ (Vs-)\)$

Frequency, Ouput Voltage and Load Curves






5.Panel Introduction



No.	Function
1	Power switch
2	Power indicator
3	Power
4	Driving output
(5)	Overcurrent indicator
6	Analog input

6.Drawings





7.Notes and suggestions

- ▶ E53.A cannot be used to drive inductive loads. If the inductive loads are driven, the product may be damaged.
- ▶ If there is no need, please do not twist the potentiometer easily.



Harbin Core Tomorrow Science & Technology Co., Ltd.

Tel: +86-451-86268790

Email: info@coremorrow.com
Website: www.coremorrow.com

Address: Building I2, No.191 Xuefu Road, Nangang District, Harbin, HLJ, China

CoreMorrow Official and CTO WeChat are below:



